



Data Collection Worksheet

Please Note: The Data Collection Worksheet (DCW) is a tool to aid integration of a PhenX protocol into a study. The PhenX DCW is not designed to be a data collection instrument. Investigators will need to decide the best way to collect data for the PhenX protocol in their study. Variables captured in the DCW, along with variable names and unique PhenX variable identifiers, are included in the PhenX Data Dictionary (DD) files.

Serum or plasma collected by standard procedures is acceptable. At least 3 mL should be provided, to have enough sample volume for a repeat analysis if needed. Samples should be stored at -20° C.

Liquid chromatography - tandem mass spectrometry (LC-MS/MS) is the preferred method to accurately measure cotinine in serum or plasma samples, especially when used to assess secondhand smoke exposure. (See source references for details.) The limit of quantitation is 0.02 ng/mL.

Protocol source: <https://www.phenxtoolkit.org/protocols/view/91706>